

VOP-026 Šternberk, s.p.

locality Vyškov

V. Nejedlého 691, 682 03 Vyškov, Czech Republic

CERTIFICATE

N° VTÚPV - 234 / 2011 / ZAHR

Applicant: EASTERN TECHNOLOGY (ZHANG ZHOU) CO., LTD.
Jinfeng Industrial District, Zhang Zhou, Fujian P.C. 363000

Product: Analog Multimeter

Model: EM365A

Manufacturer: ZHANG ZHOU EASTERN INTELLIGENT METER CO., LTD.
Zhangzhou High & New Technology Center Building, Gangqiao,
Zhangzhou City, Fujian 363000, China



Rating and principal characteristics: CAT II; 500 V
Battery Supply: 3 V DC

Test results are described in the Test Reports No.:
MTE/DAL/E11091105 (tests made by Most Technology Service Co., Limited)

The sample of tested product conforms with the requirements of the following standards harmonized with the EMC Directive 2004/108/EC

- EN 61326-1: 2006
- EN 61326-2-2: 2006

This certificate is valid until: 23. 09. 2014

After preparation of the necessary technical documentation as well as the conformity declaration, the required CE marking can be affixed on the product. Other relevant directives have to be observed. The CE mark shall appear according to this sample:



Vyškov 23. 09. 2011

Tel./Fax: +420 517 303 603
e-mail: bezdek.m@vop.cz
<http://www.vop.cz>

M. Bezděk
Milan Bezděk
Certification Head