

# C E R T I F I C A T E



of Conformity  
Low Voltage Directive 73/23/EEC  
as last amended by EEC Directive 93/68/EEC

Registration No.: AN 50017933 0001

Report No.: 15001596 031

Holder: Precision Mastech Enterprises Co.  
Room 1709, Hewlett Centre  
52 Hoi Yuen Road  
Kwun Tong, Kowloon  
Hong Kong

Product: Widerstandsmessgerät  
(Megger)

Identification: 1000V, 1000M, 2500V, 2500M, 2500V, 5000M, 25000V, 10000M  
(refer to Report 14001596 001 for detailed list)

Serial no. : Engineering Sample  
Tested acc. to : EN 61010-1:2001  
EN 61010-2-31:2002

This certificate of conformity is based on an evaluation of a sample of the above mentioned product. Technical Report and documentation are at the Licence Holder's disposal. This is to certify that the tested sample is in conformity with all revision of Annex I of Council Directive 73/23/EEC, in its latest amended version, referred to as the Low Voltage Directive. This certificate does not imply assessment of the series-production of the product and does not permit the use of a TÜV Rheinland mark of conformity. The holder of the certificate is authorized to use this certificate in connection with the EC declaration of conformity according to Annex III of the Directive.



Certification Body

Cologne, 21.05.2013

Dipl.-Ing. F. Nispel

TÜV Rheinland Product Safety GmbH - Am Grauen Stein - D-51105 Köln

CE... The CE marking may be used if all relevant and effective EC Directives are complied with.

CE The CE marking may only be used if all relevant and effective EC Directives are complied with. CE